

EDI™ Fecal *Giardia lamblia* Antigen CLIA Kit

Chemiluminescence Immunoassay (CLIA) for the detection of *Giardia lamblia* Antigen in Feces.

REF CL0813 CE IVD    

INTENDED USE

This Chemiluminescence Immunoassay (CLIA) kit is intended for the quantitative detection of *Giardia lamblia* antigen in feces using the ECL100 or ECL25. This assay is a useful tool in the diagnosis of active *Giardia lamblia* infection in acute or chronic gastroenteritis.

It is for in vitro diagnostic use only.

SUMMARY OF PHYSIOLOGY

Giardia lamblia (also known as *Giardia intestinales*) has a characteristic tear-drop shape and measures 10-15 µm in length. It has twin nuclei and an adhesive disk which is a rigid structure reinforced by supelicular microtubules. There are two median bodies of unknown function, but their shape is important for differentiating between species. There are four pairs of flagella, one anterior pair, two posterior pairs, and a caudal pair. These organisms have no mitochondria, endoplasmic reticulum, golgi, or lysosomes. *Giardia* has a two-stage life cycle consisting of trophozoite and cyst. The life cycle begins with ingested cysts, which release trophozoites (10-20 µm x 5-15 µm) in the duodenum. These trophozoites attach to the surface of the intestinal epithelium using a ventral sucking disk and then reproduce by binary fission. The trigger for encystment is unclear, but the process results in the inactive environmentally resistant form of *Giardia*—a cyst (11-14 µm x 7-10 µm) that is excreted in feces.

Giardiasis is a diarrheal illness caused by *Giardia lamblia*, after ingestion of *Giardia* cysts. Once a person has been infected with *Giardia*, the parasite lives in the intestine and is passed in the stool. Millions of herms can be released in a bowel movement from an infected human or animal. *Giardia* is found in soil, food, water, or surfaces that have been contaminated with feces from infected humans or animals. Because the parasite is protected by an outer shell, it can survive outside the body and in the environment for long periods of time. Because it is spread world-wide, *Giardia lamblia* has become one of the most important causes of chronic diarrheas. *Giardia* infection can cause a variety of intestinal symptoms either acute or chronic, which include diarrhea, gas or flatulence, greasy stools that tend to float, stomach cramps, upset stomach or nausea. These symptoms may lead to weight loss and dehydration. Some people with giardiasis have no symptoms at all. Those asymptomatic cases still shed *Giardia* cysts. Generally, symptoms of giardiasis begin 1 to 2 weeks after becoming infected and they may last 2 to 6 weeks.

The method used for the diagnosis of giardiasis in the past has been the detection of *Giardia* cysts in stool by microscopy. However, specific *Giardia* antigen CLIA greatly simplifies the diagnostic procedure, and it does not require an intact organism in the test specimen.

ASSAY PRINCIPLE

This CLIA is designed, developed, and produced for the quantitative measurement of *Giardia lamblia* in fecal samples. The assay utilizes a two-site “sandwich” technique with two antibodies that bind to different epitopes of *Giardia lamblia*.

Assay calibrators, controls, or patient samples are added directly to a reaction vessel containing streptavidin coated magnetic particles. An acridinium ester conjugated antibody and a biotin conjugated antibody are added. The magnetic particles capture the biotin antibody as well as an immuno complex in the form of “magnetic particles – biotin anti-*Giardia lamblia* antibody – *Giardia lamblia* – acridinium ester anti-*Giardia lamblia* antibody”.

The materials bound to the solid phase are held in a magnetic field while unbound materials are washed away. Then, the trigger solution is added to the reaction vessel and light generated by the reaction is measured with the ECL100 or ECL-25 analyzer. The relative light units (RLU) are proportional to the concentration of *Giardia* in the sample. The amount of analyte in the sample is determined from a stored, multi-point calibration curve and reported in Units per mL concentration.

REAGENTS: PREPARATION AND STORAGE

This test kit must be stored at 2 – 8°C upon receipt. For the expiration date of the kit, refer to the label on the kit box. All components are stable until this expiration date.

1. *Giardia* Magnetic Particle Solution (L0554)

Qty: 1 x 2.8 mL (100/kit), 2 x 2.5 mL (150/kit),
3 x 2.5 mL (250/kit)
Storage: 2 – 8°C
Preparation: Ready to Use

2. Biotin anti-*Giardia* Antibody (L0555)

Qty: 1 x 8.5 mL (100/kit), 1 x 13 mL (150/kit),
1 x 20 mL (250/kit)
Storage: 2 – 8°C
Preparation: Ready to Use

3. Acridinium Ester anti-*Giardia* Antibody (L0556)

Qty: 1 x 11 mL (100/kit), 1 x 16.5 mL (150/kit),
1 x 26.5 mL (250/kit)
Storage: 2 – 8°C
Preparation: Ready to Use

4. Fecal *Giardia lamblia* Calibrators (L0563-L0564)

Liquid *Giardia lamblia* in a bovine serum albumin-based matrix with a sodium azide preservative. Refer to vials for exact concentration.

Qty: 2 x vials, 1 mL each
Storage: 2 – 8°C
Preparation: Mix by inversion or gentle vortexing.

5. Fecal Giardia lamblia Controls (L0594- L0595)

Liquid Giardia lamblia in a bovine serum albumin-based matrix with a sodium azide preservative. Refer to vials for exact concentration.

Qty: 2 x vial of 1mL each

Storage: 2 – 8°C

Preparation: Mix by inversions or gentle vortexing.

6. Concentrated Fecal Extraction Buffer (30669) (Packed Separately)

Concentrated buffer matrix with protein stabilizers and preservative which serves as a patient sample diluent containing a surfactant in phosphate-buffered saline with a non-azide preservative.

Qty: 1 x 60 mL

Storage: 2 – 8°C

Preparation: 4X Concentrate. The contents must be diluted with 180 mL distilled water and mixed well before use.

SAFETY PRECAUTIONS

The reagents must be used in a professional laboratory environment and are for in vitro diagnostic use. Source material which contains reagents of bovine serum albumin was derived in the contiguous 48 United States. It was obtained only from healthy donor animals maintained under veterinary supervision and found free of contagious diseases. Wear gloves while performing this assay and handle these reagents as if they were potentially infectious. Avoid contact with reagents containing hydrogen peroxide. Do not get in eyes, on skin, or on clothing. Do not ingest or inhale fumes. On contact, flush with copious amounts of water for at least 15 minutes. Use Good Laboratory Practices.

MATERIALS REQUIRED BUT NOT PROVIDED

The instrument only uses materials supplied by Epitepe Diagnostics, Inc. When materials available from third-party suppliers are used, Epitepe Diagnostics, Inc. takes no responsibility for the validity of results obtained. Material is available for purchase from Epitepe Diagnostics, Inc. Please contact your distributor for more information.

1. Immunoassay Analyzer (ECL100) or (ECL25)
2. Cuvettes for ECL100 (CL011) or (CL010) ECL25
3. EDI™ Wash Reagent (P-594)
4. EDI™ Trigger Solution A and B (P595A, P595B)

SPECIMEN COLLECTION AND PREPARATION

Fresh fecal sample should be collected into a stool sample collection container. It is required to collect a minimum of 1-2 mL liquid stool sample or 1- 5 grams solid sample. The collected fecal sample must be transported to the lab in a frozen condition (-20°C). If the stool sample is collected and tested in the same day, it is allowed to be stored at 2-8°C.

Patient samples need to be diluted **1:11** with patient sample diluent before being measured. A 1x working solution of concentrated fecal sample extraction buffer (30669) is suggested in the extraction of samples used with this assay.

1. Label a test tube (12x75mm) or a 2.5 mL plastic vial.
2. Add 1 mL of diluted fecal sample extraction buffer to each tube or vial.
3. Add 100 µL of liquid stool sample to the above tube.
4. With solid stool sample, take an equivalent amount (about 80-120 mg) with a spatula or a disposable inoculation loop. Vigorously mix or vortex to dissolve stool specimen in the tube.
5. Let the extracted samples sit and sediment for 15 minutes. Make sure there is not free particle on the surface of liquid supernatant. Load the tube for sample test. Alternatively,

centrifuge the extracted fecal sample at 1000 rpm (200 g) for 3 minutes before loading the tube for testing.

*Note: The supernatant **MUST** be particle free to avoid damaging the ECL100 or ECL25 instrument. If necessary, remove the supernatant into an empty tube to ensure that no particles are present.*

CALIBRATION

An active calibration curve is required for all tests. For the assay, calibration is required for the first time use of a reagent lot and every 14 days thereafter or when either kit control is out of range. Refer to appropriate system manuals for configuring calibrators.

QUALITY CONTROL

The characteristics of patient samples are simulated through controls and are critical to validate the performance of CLIA assays due to the random access format. Use of controls is left to the discretion of the user, based on good laboratory practices, requirements, and applicable laws. We suggest performing a control test once every day. Quality control results that do not fall within acceptable ranges may indicate invalid test results.

ASSAY PROCEDURE

1. Reagents from different kit lot numbers should not be combined or interchanged. Make sure that there are no air bubbles in any reagents, calibrator and control vials.
2. Reagent Preparation
 - 2.1 Remove reagent cartridges from packaging and replace the solid caps with the provided soft caps for ECL100. For ECL25, carefully remove the aluminum foil seal on each container on the cartridges and insert soft caps.
 - 2.2 For the ECL100, take out the Magnetic Particle bottle make sure to roll between hands and gently but thoroughly mix until the magnetic particle solution is homogenous. The solution should be uniform with no clumps of magnetic particles visible; this step is vital for assay performance. For ECL25, mix the magnetic beads by moving back and forth the bottom part of the cartridge at upright position. Make sure to look inside the cartridge until the solution is uniform with no clumps of magnetic particles visible and no air bubbles. Recap the bottle. Open the top soft cap of all reagent bottles, leaving only the hollow soft rubber. The reagents are now ready to be loaded into the ECL100 or ECL 25 for calibration.
3. Assay Program

The following table illustrates the protocol used by the ECL100 or ECL25 for instrument operation.

Component	Quality Control Hole (µL)	Sample Hole (µL)
Giardia lamblia Controls (L0594-L0595)	50	-
Samples	-	50
Biotin anti-Giardia lamblia Antibody (L0555)	75	75
Giardia Magnetic Particle Solution (L0554)	25	25
Incubation Period 1		
Wash the reaction cup 3 times with the wash solution.		
Acridinium Ester anti-Giardia lamblia Antibody (L0556)	100	100
Incubation Period 2		
Wash the reaction cup 3 times with the wash solution.		
Trigger Solution A (P-595)	200	200
Trigger Solution B (P-595)	200	200

The assay total incubation time is less than 25 minutes.

INTERPRETATION OF RESULTS

The chemiluminescence analyzer calculates the concentration values of the sample and the control by a standard curve (fitting method: four parameters or point-to-point) and the measured RLU. Values are compared with the range of the marked value. If it exceeds the indicated quality control range, it indicates that the test is unqualified and needs to be re-tested.

Due to methodological differences or antibody specificity, there may be deviations between the test results of reagents from different manufacturers. Therefore, direct comparisons should not be made to avoid false interpretation.

EXPECTED VALUES

Fecal *G. lamblia* antigen concentrations were measured in stool samples collected from 55 apparently healthy adults using the EDI™ Fecal *Giardia lamblia* Antigen CLIA Kit. The suggested positive cut off is 0.5 U/mL.

It is highly recommended that each laboratory should establish their own normal range for fecal *Giardia lamblia* antigen based on local populations.

LIMITATIONS OF THE PROCEDURE

1. This product is for use on the ECL100 or ECL25 Immunoanalyzer only. Refer to the appropriate system manuals and/or Help system for a specific description of installation, start-up, operation, system performance, instructions, calibration, precautions, hazards, maintenance, and troubleshooting.
2. Reagents from different lots cannot be mixed.
3. Test results from this product should not be the sole basis for clinical diagnosis.
4. If the test sample result is higher than the upper limit of the calibration curve, it is recommended to re-measure after dilution according to a certain ratio. The measurement result is recalculated according to the dilution ratio to ensure the accuracy of the result.

PERFORMANCE CHARACTERISTICS

Linearity

Linearity was determined by three replicates each of a set of 5 dilutions of the highest standard.

Standard	Theoretical Concentration (U/mL)	Experimental Concentration (U/mL)	CV (%)*	R
1	40	33.14	7%	0.996
2	80	67.53	6%	
3	160	115.97	7%	
4	240	184.93	6%	
5	320	225.83	1%	
6	400	292.17	8%	

*This CV represents the CV for the replicates of the Experimental Concentration.

Intra-Assay Precision

A low, medium, and high concentration sample was run with 8 replicates each to evaluate the precision.

Sample	Average Concentration (U/mL)	CV (%)
L	10.18	6%
M	42.10	4%
H	141.94	7%

Inter-Assay Precision

A low, medium, and high concentration sample was run 2 runs a day with 2 replicates in each run for 5 days to determine the

Sample	Average Concentration (U/mL)	CV (%)
L	7.44	12%

M	41.50	10%
H	165.82	12%

Accuracy

Accuracy was determined after calibration by running two replicates of the third standard.

Standard	Average Concentration (U/mL)	Target Value ± 15% (U/mL)
3	6.90	5.31 – 7.19

Hook Effect

There was no hook effect observed up to 400 U/mL.

Limit of Blank

The limit of blank (LoB) was determined by 60 replicates of calibrator matrix to be 0.11 U/mL.

Limit of Detection

The limit of detection (LoD) was determined by 60 replicates of low-level samples to be 0.39 U/mL.

Limit of Quantification

The limit of quantification (LoQ) was determined by 60 replicates of low-level samples to be 0.66 U/mL.

WARRANTY

This product is warranted to perform as described in its labeling and literature when used in accordance with all instructions. Epitope Diagnostics, Inc. DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, and in no event shall Epitope Diagnostics, Inc. be liable for consequential damages. Replacement of the product or refund of the purchase price is the exclusive remedy for the purchaser. This warranty gives you specific legal rights and you may have other rights, which vary from state to state.

REFERENCES

1. F. Assche G. V. (2011). Fecal biomarkers for the diagnosis and management of inflammatory bowel disease. *Gastroenterology & hepatology*, 7(6), 396–398.
2. Bardhan PK, Beltinger J, Beltinger RW, Hossain A, Mahalanabis D, Gyr K (2000). "Screening of patients with acute infectious diarrhoea: evaluation of clinical features, faecal microscopy, and faecal occult blood testing". *Scand. J. Gastroenterol.* 35 (1): 54–60. doi:10.1080/003655200750024533. PMID 10672835
3. Beg M, Singh M, Saraswat MK, Rewari BB (2002). "Occult Gastrointestinal Bleeding: Detection, Interpretation, and Evaluation" (PDF). *JACM*. 3 (2): 153–58. Archived (PDF) from the original on 2010-11-22.
4. E. Collins, Judith F.; Lieberman, David A.; Durbin, Theodore E.; Weiss, David G. (2005). "Accuracy of Screening for Fecal Occult Blood on a Single Stool Sample Obtained by Digital Rectal Examination: A Comparison with Recommended Sampling Practice". *Annals of Internal Medicine*. 142 (2): 81. doi:10.7326/0003-4819-142-2-200501180-00006. ISSN 0003-4819
5. Federici A, Giorgi Rossi P, Borgia P, Bartolozzi F, Farchi S, Gausticchi G. (2005). The immunochemical faecal occult blood test leads to higher compliance than the guaiac for colorectal cancer screening programmes: a cluster randomized controlled trial. *J Med Screen*;12(2):83-8
6. Fraser CG, Matthew CM, Mowat NA, Wilson JA, Carey FA, Steele RJ. (2006). Immunochemical testing of individuals positive for guaiac faecal occult blood test in a screening programme for colorectal cancer: an observational study. *Lancet Oncol*;7(2):127-31.
7. Quintero E (2009). "[Chemical or immunological tests for the detection of fecal occult blood in colorectal cancer screening?]. *Gastroenterol Hepatol (in Spanish)*. 32 (8): 565–76. doi:10.1016/j.gastrohep.2009.01.179. PMID 19577340.
8. Rabeneck, L., Rumble, R. B., Thompson, F., Mills, M., Oleschuk, C., Whibley, A. Lewis, N. (2012). Fecal immunochemical tests compared with guaiac fecal occult blood tests for population-

TECHNICAL ASSISTANCE AND CUSTOMER SERVICE

For technical assistance or place an order, please contact Epitope Diagnostics, Inc. at (858) 693-7877 or fax to (858) 693-7678.

This product is developed and manufactured by














Epitope Diagnostics, Inc.
7110 Carroll Road
San Diego, CA 92121, US

Please visit our website at www.epitopediagnostics.com to learn more about our products and services.

EC	REP	MDSS GmbH Schiffgraben 41, 30175 Hannover, Germany
-----------	------------	--

GLOSSARY OF SYMBOLS (EN 980/ISO 15223)

-  In Vitro Diagnostic Device
-  European Conformity
-  Lot Number
-  Catalog Number
-  Read instructions before use
-  Number of Tests
-  Store at
-  Use by
-  Keep away from heat and direct sun light
-  Manufacturer
-  Authorized Representative in Europe