

EDI™ Fecal *Giardia lamblia* Antigen CLIA Kit

Chemiluminescence Immunoassay (CLIA) for the detection of *Giardia lamblia* Antigen in Feces.

REF CL0813 CE IVD   50, 100, 250  

INTENDED USE

This Chemiluminescence Immunoassay (CLIA) kit is intended for the quantitative detection of *Giardia lamblia* antigen in feces using the ECL100 or ECL25 Immunoassay analyzer. This test is used for screening for active *Giardia lamblia* infection in acute or chronic gastroenteritis.

For in-vitro diagnostics purpose only

SUMMARY OF PHYSIOLOGY

Giardia lamblia (also known as *Giardia intestinales*) has a characteristic tear-drop shape and measures 10-15 µm in length. It has twin nuclei and an adhesive disk which is a rigid structure reinforced by supelicular microtubules. There are two median bodies of unknown function, but their shape is important for differentiating between species. There are four pairs of flagella, one anterior pair, two posterior pairs, and a caudal pair. These organisms have no mitochondria, endoplasmic reticulum, golgi, or lysosomes. *Giardia* has a two-stage life cycle consisting of trophozoite and cyst. The life cycle begins with ingested cysts, which release trophozoites (10-20 µm x 5-15 µm) in the duodenum. These trophozoites attach to the surface of the intestinal epithelium using a ventral sucking disk and then reproduce by binary fission. The trigger for encystment is unclear, but the process results in the inactive environmentally resistant form of *Giardia*—a cyst (11-14 µm x 7-10 µm) that is excreted in feces.

Giardiasis is a diarrheal illness caused by *Giardia lamblia*, after ingestion of *Giardia* cysts. Once a person has been infected with *Giardia*, the parasite lives in the intestine and is passed in the stool. Millions of herms can be released in a bowel movement from an infected human or animal. *Giardia* is found in soil, food, water, or surfaces that have been contaminated with feces from infected humans or animals. Because the parasite is protected by an outer shell, it can survive outside the body and in the environment for long periods of time. Because it is spread world-wide, *Giardia lamblia* has become one of the most important causes of chronic diarrheas. *Giardia* infection can cause a variety of intestinal symptoms either acute or chronic, which include diarrhea, gas or flatulence, greasy stools that tend to float, stomach cramps, upset stomach or nausea. These symptoms may lead to weight loss and dehydration. Some people with giardiasis have no symptoms at all. Those asymptomatic cases still shed *Giardia* cysts. Generally, symptoms of giardiasis begin 1 to 2 weeks after becoming infected and they may last 2 to 6 weeks.

The method used for the diagnosis of giardiasis in the past has been the detection of *Giardia* cysts in stool by microscopy. However, specific *Giardia* antigen CLIA greatly simplifies the diagnostic procedure, and it does not require an intact organism in the test specimen.

ASSAY PRINCIPLE

This CLIA is designed, developed, and produced for the quantitative measurement of *Giardia lamblia* in fecal samples. The assay utilizes a two-site "sandwich" technique with two antibodies that bind to different epitopes of *Giardia lamblia*.

Assay calibrators, controls, or patient samples are added directly to a reaction vessel containing streptavidin coated magnetic particles. An acridinium ester conjugated antibody and a biotin conjugated antibody are added. The magnetic particles capture the biotin antibody as well as an immuno complex in the form of "magnetic particles – biotin anti-*Giardia lamblia* antibody – *Giardia lamblia* – acridinium ester anti-*Giardia lamblia* antibody".

The materials bound to the solid phase are held in a magnetic field while unbound materials are washed away. Then, the trigger solution is added to the reaction vessel and light generated by the reaction is measured with the ECL100 or ECL25 analyzer. The relative light units (RLU) are proportional to the concentration of *Giardia* in the sample. The amount of analyte in the sample is determined from a stored, multi-point calibration curve and reported in Units per mL concentration.

REAGENTS: PREPARATION AND STORAGE

This test kit must be stored at 2 – 8°C upon receipt. For the expiration date of the kit, refer to the label on the kit box. All components are stable until this expiration date.

Standard Batch Quantity: 100/kit

1. *Giardia* Magnetic Particle Solution (L0554)

Qty: 1 x 2.3 mL (50/kit), 1 x 2.8 mL (100/kit),
1 x 7.5 mL (250/kit)
Storage: 2 – 8°C
Preparation: Ready to Use

2. Biotin anti-*Giardia* Antibody (L0555)

Qty: 1 x 4.5 mL (50/kit), 1 x 8.5 mL (100/kit),
1 x 20 mL (250/kit)
Storage: 2 – 8°C
Preparation: Ready to Use

3. Acridinium Ester anti-*Giardia* Antibody (L0556)

Qty: 1 x 6.0 mL (50/kit), 1 x 11 mL (100/kit),
1 x 26.5 mL (250/kit)
Storage: 2 – 8°C
Preparation: Ready to Use

4. Fecal *Giardia lamblia* Calibrators (L0563-L0564)

Liquid *Giardia lamblia* in a bovine serum albumin-based matrix with a sodium azide preservative. Refer to vials for exact concentration.

Qty: 2 x vials, 1 mL each
Storage: 2 – 8°C
Preparation: Mix by inversion or gentle vortexing.

5. Fecal Giardia lamblia Controls (L0594- L0595)

Liquid Giardia lamblia in a bovine serum albumin-based matrix with a sodium azide preservative. Refer to vials for exact concentration.

Qty: 2 x vial of 1mL each

Storage: 2 – 8°C

Preparation: Mix by inversions or gentle vortexing.

6. Fecal Extraction Buffer (30669) (Packed Separately)

Fecal extraction buffer matrix with protein stabilizers and preservative which serves as a ready-to-use patient sample diluent.

Qty: 1 x 180 mL

Storage: 2 – 30°C

Preparation: Ready to Use

SAFETY PRECAUTIONS

The reagents must be used in a professional laboratory environment and are for in vitro diagnostic use. Source material which contains reagents of bovine serum albumin was derived in the contiguous 48 United States. It was obtained only from healthy donor animals maintained under veterinary supervision and found free of contagious diseases. Wear gloves while performing this assay and handle these reagents as if they were potentially infectious. Avoid contact with reagents containing hydrogen peroxide. Do not get in eyes, on skin, or on clothing. Do not ingest or inhale fumes. On contact, flush with copious amounts of water for at least 15 minutes. Use Good Laboratory Practices.

MATERIALS REQUIRED BUT NOT PROVIDED

The instrument only uses materials supplied by EpiTope Diagnostics, Inc. When materials available from third-party suppliers are used, EpiTope Diagnostics, Inc. takes no responsibility for the validity of results obtained. Material is available for purchase from EpiTope Diagnostics, Inc. Please contact your distributor for more information.

1. ECL100 Immunoassay Analyzer or ECL25 Immunoassay Analyzer
2. CL011 Cuvettes (for ECL100) or CL010 Cuvettes (for ECL25)
3. Wash Reagent (P-594)
4. Trigger Solutions A and B (P-595)

SPECIMEN COLLECTION AND PREPARATION

Fresh fecal sample should be collected into a stool sample collection container. It is required to collect a minimum of 1-2 mL liquid stool sample or 1- 5 grams solid sample. The collected fecal sample must be transported to the lab in a frozen condition (-20°C). If the stool sample is collected and tested in the same day, it is allowed to be stored at 2-8°C.

Patient samples need to be diluted approximately **1:15** with **Fecal Extraction Buffer (30669)** before being measured.

1. Label a test tube (12x75mm) or a 2.5 mL plastic vial.
2. Add 1.5 mL of fecal sample extraction buffer to each tube or vial. Note: Eppendorf Repeater pipette is recommended.
3. With liquid stool sample, using a pair of scissors, cut the pipette tips at a 45-degree angle. Then, use the pipette to transfer 150 µL of the liquid stool sample into the tube or vial above.
4. With solid stool sample, take an equivalent amount (about 100-120 mg, a green pea size) with a spatula or a disposable inoculation loop. Vigorously mix or vortex to dissolve stool specimen in the tube.

5. Let the extracted samples sit and sediment for 15 minutes. Make sure there is free particle and free bubble on the surface of liquid supernatant. Load the tube for sample test. Alternatively, centrifuge the extracted fecal sample at 1000 rpm (200 g) for 3 minutes before loading the tube for testing.

Note: The supernatant MUST be particle and bubble free to avoid damaging the ECL100 or ECL25 instrument and obtain correct test result. If necessary, remove the supernatant into an empty tube to ensure that no particles are present.

CALIBRATION

An active calibration curve is required for all tests. For the assay, calibration is required for the first-time use of a reagent lot and is valid for 28 days. However, we recommend calibration every 14 days after initial calibration or when either kit control is out of range

QUALITY CONTROL

The characteristics of patient samples are simulated through controls and are critical to validate the performance of CLIA assays due to the random access format. Use of controls is left to the discretion of the user, based on good laboratory practices, requirements, and applicable laws. We suggest performing a control test once every day. Quality control results that do not fall within acceptable ranges may indicate invalid test results.

ASSAY PROCEDURE

1. Reagents from different kit lot numbers should not be combined or interchanged. Make sure that there are no air bubbles in any reagents, calibrator and control vials.
2. Reagent Preparation
 - 2.1 Remove reagent cartridges from packaging and replace the solid caps with the provided soft caps for ECL100. For ECL25, carefully remove the aluminum foil seal on each container on the cartridges.
 - 2.2 For the ECL100, take out the Magnetic Particle bottle make sure to roll between hands and gently but thoroughly mix until the magnetic particle solution is homogenous. The solution should be uniform with no clumps of magnetic particles visible; this step is vital for assay performance.
 - Note: For ECL 100, if the Magnetic Particle Solution volume is over 3 mL, it will be provided in a glass bottle. It will need to be transferred from the glass bottle to the plastic vial in the cartridge with the rest of the reagents. Make sure the Magnetic Particle Solution is mixed well before transferring.
 - 2.3 For ECL25, mix the magnetic beads by moving back and forth the bottom part of the cartridge at upright position. Make sure to look inside the cartridge until the solution is uniform with no clumps of magnetic particles visible and no air bubbles. Recap the bottle. Open the top soft cap of all reagent bottles, leaving only the hollow soft rubber.
 - 2.4 The reagents are now ready to be loaded into the ECL100 or ECL 25 for calibration.

3. Assay Program

The following table illustrates the protocol used by the ECL100 or ECL25 for instrument operation.

Component	Quality Control Hole (µL)	Sample Hole (µL)
Giardia lamblia Controls (L0594-L0595)	75	-
Samples	-	75
Biotin anti-Giardia lamblia Antibody (L0555)	75	75

Giardia Magnetic Particle Solution (L0554)	25	25
Incubation Period 1		
Wash the reaction cup 3 times with the wash solution.		
Acridinium Ester anti-Giardia lamblia Antibody (L0556)	100	100
Incubation Period 2		
Wash the reaction cup 3 times with the wash solution.		
Trigger Solution A (P-595)	200	200
Trigger Solution B (P-595)	200	200

The assay total incubation time is less than 25 minutes.

INTERPRETATION OF RESULTS

The chemiluminescence analyzer calculates the concentration values of the sample and the control by a standard curve (fitting method: four parameters or point-to-point) and the measured RLU. Values are compared with the range of the marked value. If it exceeds the indicated quality control range, it indicates that the test is unqualified and needs to be re-tested.

Due to methodological differences or antibody specificity, there may be deviations between the test results of reagents from different manufacturers. Therefore, direct comparisons should not be made to avoid false interpretation.

EXPECTED VALUES

Fecal *G. lamblia* antigen concentrations were measured in stool samples collected from 55 apparently healthy adults using the EDI™ Fecal *Giardia lamblia* Antigen CLIA Kit. The suggested positive cut off is 0.5 U/mL.

It is highly recommended that each laboratory should establish their own normal range for fecal *Giardia lamblia* antigen based on local populations.

LIMITATIONS OF THE PROCEDURE

1. This product is for use on the ECL100 or ECL25 Immunoanalyzer only. Refer to the appropriate system manuals and/or Help system for a specific description of installation, start-up, operation, system performance, instructions, calibration, precautions, hazards, maintenance, and troubleshooting.
2. Reagents from different lots cannot be mixed.
3. Test results from this product should not be the sole basis for clinical diagnosis.
4. If the test sample result is higher than the upper limit of the calibration curve, it is recommended to re-measure after dilution according to a certain ratio. The measurement result is recalculated according to the dilution ratio to ensure the accuracy of the result.

PERFORMANCE CHARACTERISTICS

Linearity

Linearity was determined by three replicates each of a set of 5 dilutions of the highest standard.

Standard	Theoretical Concentration (U/mL)	Experimental Concentration (U/mL)	CV (%)*	R
1	40	33.14	7%	0.996
2	80	67.53	6%	
3	160	115.97	7%	
4	240	184.93	6%	
5	320	225.83	1%	
6	400	292.17	8%	

*This CV represents the CV for the replicates of the Experimental Concentration.

Intra-Assay Precision

A low, medium, and high concentration sample was run with 8 replicates each to evaluate the precision.

CL0813/V16/IVD/2024-12

Sample	Average Concentration (U/mL)	CV (%)
L	10.18	6%
M	42.10	4%
H	141.94	7%

Inter-Assay Precision

A low, medium, and high concentration sample was run 2 runs a day with 2 replicates in each run for 5 days to determine the

Sample	Average Concentration (U/mL)	CV (%)
L	7.44	12%
M	41.50	10%
H	165.82	12%

Accuracy

Accuracy was determined after calibration by running two replicates of the third standard.

Standard	Average Concentration (U/mL)	Target Value ± 15% (U/mL)
3	6.90	5.31 – 7.19

Hook Effect

There was no hook effect observed up to 400 U/mL.

Limit of Blank

The limit of blank (LoB) was determined by 60 replicates of calibrator matrix to be 0.11 U/mL.

Limit of Detection

The limit of detection (LoD) was determined by 60 replicates of low-level samples to be 0.39 U/mL.

Limit of Quantification

The limit of quantification (LoQ) was determined by 60 replicates of low-level samples to be 0.66 U/mL.

WARRANTY

This product is warranted to perform as described in its labeling and literature when used in accordance with all instructions. Epitope Diagnostics, Inc. DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, and in no event shall Epitope Diagnostics, Inc. be liable for consequential damages. Replacement of the product or refund of the purchase price is the exclusive remedy for the purchaser. This warranty gives you specific legal rights and you may have other rights, which vary from state to state.

REFERENCES

- 1: Ungar BL, Yolken RH, Nash TE, Quinn TC. Enzyme-linked immunosorbent assay for the detection of Giardia lamblia in fecal specimens. J Infect Dis. 1984 Jan;149(1):90-7.
- 2: Rosenblatt JE, Sloan LM, Schneider SK. Evaluation of an enzyme-linked immunosorbent assay for the detection of Giardia lamblia in stool specimens. Diagn Microbiol Infect Dis. 1993 May-Jun;16(4):337-41.
- 3: Stibbs HH, Samadpour M, Manning JF. Enzyme immunoassay for detection of Giardia lamblia cyst antigens in formalin-fixed and unfixed human stool. J Clin Microbiol. 1988 Sep;26(9):1665-9.
- 4: Nash TE, Herrington DA, Levine MM. Usefulness of an enzyme-linked immunosorbent assay for detection of Giardia antigen in feces. J Clin Microbiol. 1987 Jul;25(7):1169-71.
- 5: Stibbs HH. Monoclonal antibody-based enzyme immunoassay for Giardia lamblia antigen in human stool. J Clin Microbiol. 1989 Nov;27(11):2582-8.
- 6: Duque-Beltran S, Nicholls-Orejuela RS, Arevalo-Jamaica A, Guerrero-Lozano R, Montenegro S, James MA. Detection of Giardia

duodenalis antigen in human fecal eluates by enzyme-linked immunosorbent assay using polyclonal antibodies. Mem Inst Oswaldo Cruz. 2002 Dec;97(8):1165-8. Epub 2003 Jan 20.

7: Esfandiari A, Swartz J, Teklehaimanot S. Clustering of giardiasis among AIDS patients in Los Angeles County. Cell Mol Biol (Noisy-le-grand). 1997 Nov;43(7):1077-83.

8: Janoff EN, Smith PD, Blaser MJ. Acute antibody responses to Giardia lamblia are depressed in patients with AIDS. J Infect Dis. 1988 Apr;157(4):798-804.

9: Cimerman S, Cimerman B, Lewi DS. Prevalence of intestinal parasitic infections in patients with acquired immunodeficiency syndrome in Brazil. Int J Infect Dis. 1999 Summer;3(4):203-6.

10: Feitosa G, Bandeira AC, Sampaio DP, Badaro R, Brites C. High prevalence of giardiasis and strongyloidiasis among HIV-infected patients in Bahia, Brazil. Braz J Infect Dis. 2001 Dec;5(6):339-44.

TECHNICAL ASSISTANCE AND CUSTOMER SERVICE

For technical assistance or place an order, please contact Epitope Diagnostics, Inc. at (858) 693-7877 or fax to (858) 693-7678.

This product is developed and manufactured by



Epitope Diagnostics, Inc.

7110 Carroll Road

San Diego, CA 92121, US

Please visit our website at www.epitopediagnostics.com to learn more about our products and services.

EC	REP	MDSS GmbH Schiffgraben 41, 30175 Hannover, Germany
-----------	------------	--

GLOSSARY OF SYMBOLS (EN 980/ISO 15223)



In Vitro
Diagnostic
Device



European
Conformity



Lot Number



Catalog Number



Read instructions
before use



Number of Tests



Store at



Use by



Keep away from
heat and direct
sun light



Manufacturer



Authorized
Representative
in Europe