

EDI™ Human Chromogranin A ELISA Kit

Enzyme Linked Immunosorbent Assay (ELISA) for the measurement of Human Chromogranin A Level in EDTA Plasma

REF KT-812 CE IVD 

INTENDED USE

This ELISA (enzyme-linked immunosorbent assay) kit is intended for the quantitative determination of human chromogranin A levels in EDTA-plasma samples. This test may be used as an aid for detecting patients with pheochromocytoma and neuroendocrine tumors (carcinoids).

For In-vitro diagnostic use only

SUMMARY OF PHYSIOLOGY

Chromogranin A is a 49 kDa acidic protein that consists of 439 amino acids encoded on chromosome 14. Chromogranin A has been identified in a number of normal and neoplastic endocrine tissues. It is demonstrated that an elevated level of circulating chromogranin A is a marker for tumors of neuroendocrine origin. However, the most significant clinical use of chromogranin A is related to the diagnostic procedure in patients with pheochromocytoma. The following is a short summary of the potential usages of chromogranin A.

1. A very sensitive (83%) and highly specific (96%) marker in the evaluation of actual or suspected pheochromocytoma. Drugs commonly employed in the diagnosis or treatment of pheochromocytoma have little effect on the plasma chromogranin A level, which is a great advantage of measuring chromogranin A over catecholamines.
2. To ascertain the source of a tumor. A high chromogranin A level indicates that the tumor arises from neuroendocrine tissues.
3. Endocrine tumors that do not produce their specific hormones, for example, calcitonin negative but chromogranin A positive, C-cell carcinoma; zero-cell carcinoma; beta-cell carcinoma; parathyroid carcinoma.

ASSAY PRINCIPLE

This ELISA is designed, developed and produced for the quantitative measurement of human chromogranin A in EDTA-plasma sample. The assay utilizes the two-site "sandwich" technique with two selected antibodies that bind to different epitopes of human chromogranin A.

Assay calibrators, controls and patient samples are added directly to wells of microplate that is coated with a polyclonal chromogranin A antibody. After the first incubation period, the antibody on the wall of microtiter well captures human chromogranin A in the sample and unbound protein in each microtiter well is washed away. Then a horseradish peroxidase (HRP)-labeled monoclonal anti-human chromogranin A antibody is added to each microtiter well and a "sandwich" of "monoclonal antibody – human chromogranin A – polyclonal antibody" is formed. The unbound monoclonal antibody is removed in the subsequent washing step. For the detection of this immunocomplex, the well is incubated with a substrate solution in a timed reaction and then measured in a spectrophotometric microplate reader. The enzymatic activity of the immunocomplex bound to the chromogranin A on the wall of the microtiter well is directly *proportional* to the amount of chromogranin A in the sample. A calibration curve is generated by plotting the absorbance versus the respective human chromogranin A concentration for each calibrator on point-to-point or cubical scales or 4 parameter curve fits. The concentration of human chromogranin A in test samples is determined directly from this calibration curve.

REAGENTS: PREPARATION AND STORAGE

This test kit must be stored at 2 – 8°C upon receipt. For the expiration date of the kit, refer to the label on the kit box. All components are stable until this expiration date.

1. Anti-CgA Antibody Coated Microplate (30063)

Microplate coated with human chromogranin A antibody.

Qty: 1 x 96 well microplate

Storage: 2 – 8°C

Preparation: Ready to use

2. CgA Tracer Antibody (30153)

HRP-labeled anti-human CgA antibody in a stabilized protein matrix.

Qty: 1 x 0.6 mL

Storage: 2 – 8°C

Preparation: 21X Concentrated. Must be diluted with tracer antibody diluent (30017) and mixed well before use.

3. Tracer Antibody Diluent (30017)

Buffer for tracer antibody dilution according to the assay procedures.

Qty: 1 x 12.0 mL

Storage: 2 – 8°C

Preparation: Ready to use

4. CgA Assay Buffer (30074)

Phosphate buffered saline with bovine serum albumin.

Qty: 1 x 30.0 mL

Storage: 2 – 8°C

Preparation: Ready to use

5. ELISA Wash Concentrate (10010)

Surfactant in phosphate buffered saline with non-azide preservative.

Qty: 1 x 30.0 mL

Storage: 2 – 25°C

Preparation: 30X Concentrated. Must be diluted with 870 mL distilled water and mixed well before use.

6. ELISA HRP Substrate (10020)

Tetramethylbenzidine (TMB) with stabilized hydrogen peroxide.

Qty: 1 x 12.0 mL

Storage: 2 – 8°C

Preparation: Ready to use

7. ELISA Stop Solution (10030)

0.5 M sulfuric acid.

Qty: 1 x 12.0 mL

Storage: 2 – 25°C

Preparation: Ready to use

8. Chromogranin A Calibrators Levels 1 to 5 (30064-30068)

Human chromogranin A in a lyophilized bovine serum albumin-based matrix with a ProClin preservative.

Qty: 5 x Vials

Storage: 2 – 8°C (Lyophilized), <-20°C (Reconstituted)

Do not exceed 3 freeze-thaw cycles.

Preparation: Must be reconstituted with 0.5 mL of demineralized water, allowed to sit for 10 minutes, and mix microwell by inversions or gentle vortexing. Make sure that all solids are dissolved completely prior to use.

9. Chromogranin A Controls (30069-30070)

Human chromogranin A in a lyophilized bovine serum albumin-based matrix with a ProClin preservative.

Qty: 2 x Vials

Storage: 2 – 8°C (Lyophilized), <-20°C (Reconstituted)

Do not exceed 3 freeze-thaw cycles.

Preparation: Must be reconstituted with 0.5 mL of demineralized water, allowed to sit for 10 minutes, and mix microwell by inversions or gentle vortexing. Make sure that all solids are dissolved completely prior to use.

SAFETY PRECAUTIONS

The reagents must be used in a professional laboratory environment and are for in vitro diagnostic use. Source material which contains reagents of bovine serum albumin was derived in the contiguous 48 United States. It was obtained only from healthy donor animals maintained under veterinary supervision and found free of contagious diseases. Wear gloves while performing this assay and handle these reagents as if they were potentially infectious. Avoid contact with reagents containing hydrogen peroxide. Do not get in eyes, on skin, or on clothing. Do not ingest or inhale fumes. On contact, flush with copious amounts of water for at least 15 minutes. Exercise Good Laboratory Practices.

MATERIALS REQUIRED BUT NOT PROVIDED

1. Precision single channel pipettes capable of delivering 25 µL, 50 µL, 100 µL, and 1000 µL
2. Repeating dispenser suitable for delivering 100 µL
3. Disposable pipette tips suitable for above volume dispensing
4. Disposable 12 x 75 mm or 13 x 100 glass or plastic tubes
5. Disposable plastic 100 mL and 1000 mL bottles with caps
6. Aluminum foil
7. Deionized or distilled water
8. Plastic microtiter well cover or polyethylene film
9. ELISA multichannel wash bottle or automatic (semi-automatic) washing system
10. Spectrophotometric microplate reader capable of reading absorbance at 620 nm
11. Calibrated timer

SPECIMEN COLLECTION & STORAGE

Only 50 µL of human EDTA-plasma is required for human chromogranin A measurement in a duplicate. No special preparation of individual is necessary prior to specimen collection. Whole blood should be collected with lavender-top Vacutainer. Separate the plasma from cells by centrifugation at 850 – 1500 xg for 10 minutes. The plasma should be separated from the cells within one hour of blood collection and transferred to a clean test tube. Plasma samples should be stored below –20°C if the assay is not to be performed within 72 hours. Otherwise, the plasma samples can be stored at room temperature for up to 72 hours. It is important that the plasma samples should not be stored at 2 – 8°C in any circumstance. Avoid more than three freeze-thaw cycles of specimen.

Intact chromogranin A molecule shows better stability in EDTA-plasma than serum, it is recommended not to use serum sample for measurement.

Collected EDTA-plasma samples should be shipped to designated laboratory in a frozen condition with dry ice. In case frozen condition is unavailable, samples should be shipped at room temperature in an insulated container for maximum 48 hour delivery. Samples should not be shipped with a blue ice pack or refrigerated.

ASSAY PROCEDURE

1. Reagent Preparation

1. Prior to use allow all reagents to come to room temperature (20-25 °C). Reagents from different kit lot numbers should not be combined or interchanged.
2. ELISA Wash Concentrate (10010) must be diluted to working solution prior use. Please see REAGENTS section for details.
3. Reconstitute all assay calibrators (30064-30068) and controls (30069, 30070) by adding 0.5 mL of demineralized water to each vial. Allow the calibrators and controls to sit undisturbed for 10 minutes, and then mix well by inversions or gentle vortexing. Make sure that all solids are dissolved completely prior to use. These reconstituted calibrators and controls must be stored at -20°C or below. Do not exceed 3 freeze-thaw cycles.

2. Manual Assay Procedure

1. Place a sufficient number of Anti-CgA Antibody Coated microwell strips (30063) in a holder to run calibrators (30064 - 30068), controls (30069, 30070), and samples in duplicate.
2. Test Configuration:

Row	Strip 1	Strip 2	Strip 3
A	Calibrator Level 1	Calibrator Level 5	Sample 2
B	Calibrator Level 1	Calibrator Level 5	Sample 2
C	Calibrator Level 2	Control 1	Sample 3
D	Calibrator Level 2	Control 1	Sample 3
E	Calibrator Level 3	Control 2	Sample 4
F	Calibrator Level 3	Control 2	Sample 4
G	Calibrator Level 4	Sample 1	Sample 5
H	Calibrator Level 4	Sample 1	Sample 5

3. Add **25 µL** of calibrators (30064 - 30068), controls (30069, 30070), and samples into the designated microwells.
4. Add **100 µL** of assay buffer (30074) into each microwell.
5. Cover the plate with one plate sealer and aluminum foil. Incubate at **room temperature (20-25 °C)** and shaking at **170 rpm** on an orbital shaker or **450 rpm** on an ELISA plate shaker for **120 minutes**.
6. Remove the plate sealer. Aspirate the contents of each well. Wash each well **5 times** by dispensing **350 µL** of diluted wash solution (10010) into each well, then completely aspirating the contents. Alternatively, an automated microplate washer can be used.
7. Prepare the antibody working solution by 1:21 fold dilution of the tracer antibody (30153) with the diluent (30017). For each strip, it is required to mix 1 mL of the tracer antibody diluent with 50 µL of the tracer antibody in a clean test tube.
Note: This antibody working solution should be freshly prepared.
8. Add **100 µL** of antibody working solution to each well. Mix by gently tapping the plate.
9. Cover the plate with one plate sealer and aluminum foil. Incubate at **room temperature (20-25 °C)** and shaking at **170 rpm** on an orbital shaker or **450 rpm** on an ELISA plate shaker for **60 minutes**.
10. Remove the plate sealer. Aspirate the contents of each well. Wash each well **5 times** by dispensing **350 µL** of diluted wash solution (10010) into each well, then completely aspirating the contents. Alternatively, an automated microplate washer can be used.
11. Add **100 µL** of ELISA HRP Substrate (10020) into each of the wells. Mix by gently tapping the plate.
12. Cover the plate with one plate sealer and aluminum foil. Incubate at **room temperature (20-25 °C)** for **20 minutes**.
13. Remove the aluminum foil and plate sealer. Add **100 µL** of ELISA Stop Solution (10030) into each of the wells. Mix gently by tapping the plate.
14. Read the absorbance at **450 nm** within **10 minutes** with a microplate reader.

Note: to reduce the background, one can set the instrument to dual wavelength measurement at 450 nm with background wavelength correction set at 595 nm, 620 nm or 630 nm.

3. Automated Assay Procedure (Dynex DS-2)

1. Load a sufficient number of Anti-CgA Antibody Coated microwell strips (30063) in a holder to run calibrators (30064 - 30068), controls (30069, 30070), and samples in duplicate.
2. Load sufficient CgA tracer antibody (30153).
3. Prepare and load kit calibrators (30064 – 30068), controls (30069, 30070), patient samples, substrate (10020), stop solution (10030), diluted wash buffer (10010) onto the system accordingly.
4. Add **25 µL** of calibrators, controls and patient samples into the designated microwells.
5. Add **100 µL** of assay buffer (30074) to each well.

6. Incubate plate with shaking and at **room temperature (20-25 °C)** for **100 to 120 minutes**.
7. Wash each well **4 - 5 times**.
8. Prepare the antibody working solution by 1:21 fold dilution of the tracer antibody (30153) with the diluent (30017). For each strip, it is required to mix 1 mL of the tracer antibody diluent with 50 µL of the tracer antibody in a clean test tube.
Note: This antibody working solution should be freshly prepared.
9. Add **100 µL** of antibody working solution to each well.
10. Incubate plate with shaking and at **room temperature (20-25 °C)** for **40 – 60 minutes**.
11. Wash each well **4 - 5 times**.
12. Add **100 µL** of ELISA HRP Substrate (10020) into each of the wells.
13. Incubate plate at **room temperature (20-25 °C)** for **15 - 20 minutes**.
14. Add **100 µL** of ELISA Stop Solution (10030) into each of the wells. Mix gently.
15. Read the absorbance at **450 nm or 450/620 nm**.
Note: Open automated ELISA system other than DS-2 can also be used. This may require adjustment of the protocol. It is very important to incubate the assay 18-22°C. A change of incubation temperature would cause unsatisfactory calibration curve and erroneous test results.

PROCEDURAL NOTES

1. It is recommended that all samples be assayed in duplicate. The average absorbance reading of each duplicate should be used for data reduction and the calculation of results.
2. If a Tecan is used for pipetting, it is recommended to add 200 µL assay buffer before adding the 15 µL assay calibrators, controls and test samples into each designated well. This is the same as the procedure with DS-2, but a reverse with the manual procedure.
3. Keep light-sensitive reagents in the original amber bottles.
4. Store any unused antibody-coated strips in the foil zipper bag with desiccant to protect from moisture.
5. Careful technique and use of properly calibrated pipetting devices are necessary to ensure reproducibility of the test.
6. Incubation times or temperatures other than those stated in this insert may affect the results.
7. Avoid air bubbles in the microwell as this could result in lower binding efficiency and higher CV% of duplicate reading.
8. All reagents should be mixed gently and thoroughly prior use. Avoid foaming.
9. We strongly recommend using 4-Parameter curve fit for control and patient sample calculation. Other curve fit programs such as Point-to-Point, Log-Log, Log-Linear, etc. may give a poor linear recovery.

INTERPRETION OF RESULTS

The human chromogranin A concentrations for the controls and patient samples are read directly from the calibration curve using their respective corrected absorbance.

Laboratory should report test results directly derived from the assay. For samples showing a higher than 90% value of the highest assay calibrator, it is strongly recommended that the patient sample is diluted 1:100 with assay buffer and re-assay the diluted sample for a more accurate test result. For example, the highest assay calibrator is about 550 ng/mL, any sample that shows a value greater than 500 ng/mL (90% of 550 ng/mL) should be repeated with 1:100 diluted sample. If the 1:100 diluted samples still shows a higher value than that of the highest assay calibrator, one can either report the sample value as greater than the highest assay calibrator (e.g. > 56,000 ng/mL) or further measure 1:10,000 diluted sample. It is preferred to obtain a diluted sample value located between calibrator level 2 and level 4, wherein, this measured value is multiplied by the dilution factor to obtain the report value for the patient.

LIMITATIONS OF THE PROCEDURE

1. Since there is no Gold Standard concentration available for human chromogranin A measurement, the values of assay calibrators were established by correlation to a highly purified chromogranin A calibrator.
2. For sample values reading greater than the highest calibrator or 90% value of the highest calibrator, it is recommended to re-perform the assay with diluted samples.
3. Storing samples at refrigerated condition causes significant degradation of intact chromogranin A into small fragments. These fragments may cause interference of the assay resulting in false low test result.
4. Serum samples are not as stable as EDTA-plasma samples. Therefore, it is strongly recommended to use EDTA-plasma sample for chromogranin A measurement.
5. Bacterial or fungal contamination of plasma specimens or reagents, or cross-contamination between reagents may cause erroneous results.
6. Water deionized with polyester resins may inactive the horseradish peroxidase enzyme.

QUALITY CONTROL

To assure the validity of the results each assay should include adequate controls with known chromogranin A levels. We recommend that all assays include the laboratory's own chromogranin A controls in addition to those provided with this kit.


EXPECTED VALUES

Seventy-two normal adult sera were measured with this human chromogranin A ELISA. The normal range was found to be less than 100 ng/mL. Five patients with pheochromocytoma showed a chromogranin A level of significantly over 100 ng/mL and one of them reached 400,000 ng/mL. It is highly recommended that each laboratory establish its own normal cut-off level. Paired EDTA-Plasma and Serum samples give almost the same values. Although a chromogranin A level above 100 ng/mL would be an aid in clinical diagnosis, it is recommended to establish a baseline level of chromogranin A for each patient in order to monitor cancer patients after surgery, especially if this assay is used for the monitoring of prostate cancer patients. A clear surge of chromogranin A level would indicate an increased cancer cell activity.

EXAMPLE DATA

A typical absorbance data and the resulting calibration curve from human chromogranin A ELISA are represented in a below table.

Well ID	Absorbance readings at 450/620 nm			Concentration (ng/mL)
	Readings	Average	Corrected	
Calibrator Level 1: 0 ng/mL	0.078	0.0760	0.000	
	0.074			
Calibrator Level 2: 31 ng/mL	0.242	0.240	0.164	
	0.239			
Calibrator Level 3: 93 ng/mL	0.757	0.748	0.672	
	0.740			
Calibrator Level 4: 280 ng/mL	1.840	1.802	1.726	
	1.763			
Calibrator Level 5: 830 ng/mL	3.139	3.059	2.983	
	2.980			
Control 1	0.444	0.447	0.371	
	0.451			
Control 2	1.300	1.276	1.200	
	1.252			



Note: This curve should not be used in lieu of calibrator curve run with each assay.

PERFORMANCE CHARACTERISTICS

Sensitivity

The sensitivity of the human chromogranin A ELISA was determined by the 95% confidence limit on 20 duplicates of a zero standard and was found to be ~5 ng/mL.

Hook Effect

This assay has showed that it did not exhibit any high dose "hook" effect up to 1,000,000 ng/mL.

Reproducibility and Precision

The intra-assay precision was validated by measuring two control samples in a single assay with 8 replicate determinations and the inter-assay reproducibility was validated by measuring two control samples in duplicate in 12 individual assays. The results are summarized below:

Sample No.	Intra-assay Precision		Inter-assay Reproducibility	
	Mean (ng/mL)	CV (%)	Mean (ng/mL)	CV (%)
1	63.5 ng/mL	4.2%	61.9 ng/mL	6.7%
2	209 ng/mL	3.6%	213.3 ng/mL	5.6%

Linearity

Three human samples were diluted with assay buffer and assayed. The results are summarized below:

Samples	Observed Concentration (ng/mL)	Expected Concentration (ng/mL)	Recovery (%)
Sample A	285.0 ng/mL	-	-
50.0%	138.0 ng/mL	143.0 ng/mL	96%
25.0%	75.0 ng/mL	72.0 ng/mL	104%
12.5%	37.9 ng/mL	36.0 ng/mL	105%
6.25%	19.5 ng/mL	18.0 ng/mL	108%
Sample B	61.8 ng/mL	-	-
50.0%	32.1 ng/mL	30.9 ng/mL	104%
25.0%	15.9 ng/mL	15.5 ng/mL	103%
12.5%	7.2 ng/mL	7.7 ng/mL	94%

Spike Recovery

Two patient samples were spiked with various amounts of human chromogranin A control samples and assayed. The results are summarized below:

Sample No.	Original Value	Amount Spiked	Observed Concentration (ng/mL)	Expected Concentration (ng/mL)	Recovery (%)
1	62.8	31	45.2	46.9	96%
		93	75.6	77.9	97%
		280	152.8	171.4	89%
2	289.0	31	152.2	160.0	95%
		93	176.0	191.0	92%
		280	288.2	284.5	101%

WARRANTY

This product is warranted to perform as described in its labeling and literature when used in accordance with all instructions. Epitope Diagnostics, Inc. DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, and in no event shall Epitope Diagnostics, Inc. be liable for consequential damages. Replacement of the product or refund of the purchase price is the exclusive remedy for the purchaser. This warranty gives you specific legal rights and you may have other rights, which vary from state to state.

REFERENCES

1. Pirker RA, Pont J, Pöhl R, Schütz W, Griesmacher A, Müller MM. Usefulness of chromogranin A as a marker for detection of relapses of carcinoid tumours. Clin Chem Lab Med 1998; 36:837-40.
2. Kimura N, Miura W, Noshiro T, Mizunashi K, Hanew K, Shimizu K, et al. Plasma chromogranin A in pheochromocytoma, primary hyperparathyroidism and pituitary adenoma in comparison with catecholamine, parathyroid hormone and pituitary hormones. Endocr J 1997; 44:319-27.
3. Hendy GN, Bevan S, Mattei MG, Moulard AJ. Chromogranin A. Clin Invest Med 1995; 18:47-65.
4. Deftos LJ. Chromogranin A: its role in endocrine function and as an endocrine and neuroendocrine tumor marker. Endocrine Reviews: 1991;12:181-7.
5. Sobol RE, Memoli V, Deftos LJ. Hormone-negative, chromogranin A-positive endocrine tumors. N Engl J Med 1989; 320:444-7.

TECHNICAL ASSISTANCE AND CUSTOMER SERVICE

For technical assistance or to place an order, please contact Epitope Diagnostics, Inc. at +1 (858) 693-7877 or fax to +1 (858) 693-7678 or email at cs@epitopediagnostics.com

This product is developed and manufactured by

Epitope Diagnostics, Inc.

7110 Carroll Road

San Diego, CA 92121, USA

Please visit our website at www.epitopediagnostics.com to learn more about our products and services.



MDSS GmbH
Schiffgraben 41,
30175 Hannover, Germany

GLOSSARY OF SYMBOLS (EN 980/ISO 15223)



In Vitro Diagnostic Device



European Conformity



Lot Number



Catalog Number



Read Instructions before Use



Number of Tests



Store at



Use by



Keep Away from Heat and Direct Sun light



Manufacturer



Authorized Representative in Europe

SHORT ASSAY PROCEDURE

1. Manual Assay Procedure

1. Add **25 µL** of the calibrators, controls, and samples into the designated microwells.
2. Add **100 µL** of the assay buffer to each well.
3. Mix, cover, and incubate at **room temperature (20-25 °C)** and shaking at **170 rpm** on an orbital shaker or **450 rpm** on an ELISA plate shaker for **120 minutes**.
4. Wash each well five times.
5. Add **100 µL** of the tracer antibody to each well.
6. Cover and incubate at **room temperature (20-25 °C)** with **orbital shaking at 170 or 450 rpm** for **60 minutes**.
7. Wash each well five times.
8. Add **100 µL** of substrate to each well.
9. Cover and incubate at **room temperature (20-25 °C)** for **20 minutes**.
10. Add **100 µL** of the stop solution to each well.
11. Read the absorbance at **450 nm**.

2. Automated Assay Procedure

1. Prepare and load reagents.
2. Add **25 µL** of calibrators, controls and patient samples into the designated microwells.
3. Add **100 µL** of assay buffer (30074) to each well.
4. Incubate plate with shaking and at **room temperature (20-25 °C)** for **100 to 120 minutes**.
5. Wash each well **4 - 5 times**.
6. Add **100 µL** of *antibody working solution* to each well.
7. Incubate plate with shaking and at **room temperature (20-25 °C)** for **40 – 60 minutes**.
8. Wash each well **4 - 5 times**.
9. Add **100 µL** of ELISA HRP Substrate into each of the wells.
10. Incubate plate at **room temperature (20-25 °C)** for **15 - 20 minutes**.
11. Add **100 µL** of ELISA Stop Solution into each of the wells. Mix gently.
12. Read the absorbance at **450 nm** or **450/620 nm**.