

EDI™ Human VDBP (GC Globulin) ELISA Kit

Enzyme Linked ImmunoSorbent Assay (ELISA) for the Quantitative Measurement of Vitamin D Binding Protein Level in Serum or Plasma



I. INTENDED USE

This test kit is intended for use in the quantitative determination of free and not actin complex bound Vitamin D Binding Protein (VDBP, also known as GC Globulin) in human serum and plasma. This kit is for in vitro diagnostic use only.

II. PHYSIOLOGY

Vitamin D-binding protein (DBP) is a 58 kDa circulating glycoprotein secreted by the liver. It binds about 85% to 90% of circulating 25-OH-D_{2/3} and 1,25-(OH)₂-D₃, regulates the bioavailability of active vitamin D, and transports them to target organs. There is only less than one percent 25-OH-D_{2/3} is in the free form in the circulation. The full length DBP contains 476-amino acids, including a 16-amino acid signal sequence. Biologically, DBP may involve directly and indirectly in regulating bone metabolism. Circulating DBP also binds to actin at 1:1 molecule ratio, while this protein complex is removed by kidney. In patient with trauma, sepsis or multiple organ failure, DBP concentration decreases. Urine DBP is reported to be a biomarker of major renal event in patient undergoing coronary angiography.

III. ASSAY PRINCIPLE

The quantitative VDBP ELISA is a solid phase **competitive** immunoassay designed to detect VDBP. Microwells are coated with anti-VDBP antibody. Assay calibrators, control and diluted unknown serum or plasma specimens are added to the microwells along with GC Globulin HRP conjugated antibody. After an incubation period, the immunocomplex of solid phase bound "VDBP Antibody-VDBP" is formed, which inhibits the formation of "VDBP Antibody – HRP Conjugated VDBP". Unbound HRP Conjugated VDBP are removed with a washing step. During a second incubation with TMB substrate, a blue color is developed. The enzyme-substrate reaction is stopped by the addition of sulfuric acid. The absorbance of assay calibrators, controls and unknown specimens are measured by an ELISA plate reader with wavelength set at 450 nm. The color intensity is inversely proportional to the amount of VDBP present in the calibrators controls and specimens.

IV. REAGENTS: Preparation and Storage

This test kit must be stored at 2 – 8 °C upon receipt. For the expiration date of the kit refer to the label on the kit box. All components are stable until this expiration date.

Prior to use allow all reagents to come to room temperature. Reagents from different kit lot numbers should not be combined or interchanged.

1. VDBP Antibody Coated Microplate (Cat. No. 30974)

One microplate with twelve by eight well-breakable strips (96 wells total) coated with anti-VDBP antibody. The plate is framed and sealed in a foil zipper bag with a desiccant. This reagent should be stored at 2-8°C and is stable until the expiration date on the kit box.

2. HRP Conjugated VDBP (Cat. No. 30975)

One bottle containing 12 mL ready to use HRP labeled GC Globulin in a stabilized protein matrix. This reagent should be stored in 2-8°C and is stable until the expiration date on the kit box.

4. VDBP Concentrated Sample Diluent (Cat. No. 30976)

One bottle containing 30 mL of 6-fold concentrated buffer matrix with protein stabilizers and preservative. This reagent should be stored at 2 – 8 °C and is stable until the expiration date on the kit box. Before use the concentrated buffer must be diluted by mixing with 150 mL of demineralized water or distilled water to make the 1x VDBP Sample Dilution Buffer.

5. ELISA Wash Concentrate (Cat. No. 10010)

One bottle containing 30 mL of 30-fold concentrate. Before use, the contents must be diluted with 870 mL of demineralized water and mixed well. Upon dilution, this yields a working wash solution containing a surfactant in phosphate buffered saline with a non-azide, non-mercury preservative. The diluted wash solution may be stored at room temperature and is stable until the expiration date on the kit box.

6. ELISA HRP Substrate (Cat. No. 10020)

One bottle containing 15 mL of tetramethylbenzidine (TMB) with hydrogen peroxide. This reagent should be stored at 2-8°C and is stable until the expiration date on the kit box.

7. ELISA Stop Solution (Cat. No. 10030)

One bottle containing 15 mL of stop solution. This reagent may be stored at 2-8°C or room temperature and is stable until the expiration date on the kit box.

8. VDBP Calibrators 1 to 6 (Cat. No. 30981 - 30986)

Six vials contain 0.5 mL each of liquid GC Globulin in a bovine serum albumin-based matrix with a non-azide preservative. **Refer to the vial for exact concentration.** This reagent should be stored in 2-8°C and is stable until the expiration date on the kit box.

9. VDBP Controls (Cat. No. 30986, 30987)

Two vials each contain 0.5 mL of liquid GC Globulin in a bovine serum albumin-based matrix with a non-azide preservative. **Refer to vials for concentration range for each control.** This reagent should be stored in 2-8°C and is stable until the expiration date on the kit box.

V. SAFETY PRECAUTIONS

The reagents must be used in professional laboratory. Source material for reagents containing bovine serum was derived in the contiguous 48 United States. It was obtained only from healthy donor animals maintained under veterinary supervision and found free of contagious diseases. Wear gloves while performing this assay and handle these reagents as if they are potentially infectious. Avoid contact with reagents containing TMB, hydrogen peroxide, or sulfuric acid. TMB may cause irritation to skin and mucous membranes and

cause an allergic skin reaction. TMB is a suspected carcinogen. Sulfuric acid may cause severe irritation on contact with skin. Do not get in eyes, on skin, or on clothing. Do not ingest or inhale fumes. On contact, flush with copious amounts of water for at least 15 minutes. Use Good Laboratory Practices.

VI. MATERIALS REQUIRED BUT NOT PROVIDED

1. Precision single channel pipettes capable of delivering 25 μ l, 100 μ l, 500 μ L, etc.
2. Disposable pipette tips suitable for above volume dispensing.
3. DT-water or DI-water
4. Aluminum foil.
5. Plastic microtiter well cover or polyethylene film.
6. ELISA multichannel wash bottle or automatic (semi-automatic) washing system.
7. Spectrophotometric microplate reader capable of reading absorbance at 450 nm.
8. Oven or water bath with adjustable temperature to 37°C.

VII. SPECIMEN COLLECTION

Both serum and EDTA-plasma can be used with this test kit. Only 10 μ L total of human EDTA-plasma or serum is required for the VDBP measurement in duplicate. No special preparation of individual is necessary prior to specimen collection. Collect whole blood with Vacutainer and separate the serum or plasma from cells according to manufacturer's instruction. Sample can be kept at - 15°C. Avoid more than three freeze-thaw cycles of specimen.

VIII. ASSAY PREPARATION

1. Reagent Preparation

- (1) Prior to use allow all reagents to come to room temperature. Reagents from different kit lot numbers should not be combined or interchanged.
- (2) ELISA Wash Concentrate and sample diluent must be diluted to working solution prior to use. Please see REAGENTS section for details.
- (3) Each unknown sample needs to be diluted 1:200 using 1x VDBP Sample Dilution Buffer as a sample diluent.
 - Mark 13 x 73 mm tubes accordingly
 - Add 2 mL of VDBP Sample Dilution Buffer into each plastic tube.
 - Add 10 μ L unknown specimen to the designated tube and mix well.

IX. ASSAY PROCEDURE

Test Configuration

ROW	STRIP 1	STRIP 2	STRIP 3	STRIP 4
A	CAL 1	CAL 5	SAMPLE 1	SAMPLE 5
B	CAL 1	CAL 5	SAMPLE 1	SAMPLE 5
C	CAL 2	CAL 6	SAMPLE 2	SAMPLE 6
D	CAL 2	CAL 6	SAMPLE 2	SAMPLE 6
E	CAL 3	C 1	SAMPLE 3	
F	CAL 3	C 1	SAMPLE 3	
G	CAL 4	C 2	SAMPLE 4	
H	CAL 4	C 2	SAMPLE 4	

- (1) Add 15 μ l of calibrators, controls and diluted 1:200 test samples into the designated microwells.
- (2) Add 100 μ l HRP Conjugated VDBP to each well.
- (3) Mix contents of wells gently for 5 - 10 seconds. Seal the plate securely, cover with aluminum foil, and incubate at **room temperature for 60 minutes.**
- (4) Wash each well 5 times by dispensing 350 μ L of working wash solution into each well and then completely aspirating the contents. Alternatively, an automated microplate washer can be used.

- (5) Add 100 μ l TMB reagent to each of the wells.
- (6) Cover plate with aluminum foil, and incubate at **room temperature for 20 minutes.**
- (7) Immediately add 100 μ L of ELISA Stop Solution into each of the wells. Mix gently.
- (8) Read the absorbance at 450 nm. It is recommended to use 4-parameter curve fit to calculate the sample results.

X. PROCEDURAL NOTES

1. It is recommended that all calibrators and unknown samples be assayed in duplicate. The average absorbance reading of each duplicate should be used for data reduction and the calculation of results. It is recommended to add external controls to each assay.
2. Keep light sensitive reagents in the original amber bottles.
3. Store any unused antibody coated strips in the foil zipper bag with desiccant to protect from moisture.
4. Careful technique and use of properly calibrated pipetting devices are necessary to ensure reproducibility of the test.
5. Incubation times or temperatures other than those stated in this insert may affect the results.
6. Avoid air bubbles in the microwell as this could result in lower binding efficiency and higher CV% of duplicate reading.
7. All reagents should be mixed gently and thoroughly prior to use. Avoid foaming.
8. It is important to seal the plate properly to avoid evaporation.

XI. INTERPRETATION OF RESULTS

It is recommended to use a 4-parameter calibration curve fitting.

1. Calculate the average absorbance for each pair of duplicate test results.
2. The calibration curve is generated by the corrected absorbance of all calibration levels on the ordinate against the calibrator concentration. Appropriate computer assisted data reduction programs should be used for the calculation of results.

VDBP concentrations for the control and samples must be multiplied by the dilution factor, which is 200.


XII. EXAMPLE DATA AND STANDARD CURVE

A typical absorbance data and the resulting calibration curve from this 25-OH Vitamin D EIA is represented. **This curve should not be used in lieu of calibration curve generated with each assay.**

Well I.D	OD 450 nm Absorbance		B/B0
	Readings	Average	
Cal - 1: 0 μ g/mL	3.116	3.141	100%
	3.165		
Cal - 2: 0.5 μ g/mL	1.540	1.533	49%
	1.527		
Cal - 3: 1.3 μ g/mL	1.154	1.095	35%
	1.036		
Cal - 4: 3.2 μ g/mL	0.568	0.598	19%
	0.629		
Cal - 5: 8.0 μ g/mL	0.373	0.373	12%
	0.373		
Cal - 6: 20.1 μ g/mL	0.208	0.202	6%
	0.196		

	OD	Mean OD	Result	VDBP concentration (µg/mL)
Control 1	1.290	1.061	1.251	250.2
	1.211			
Control 2	6.614	0.422	6.473	1294.6
	6.331			

VDBP Calibrator Curve



XIII. EXPECTED VALUE

18 sera and plasma from normal human samples were measured using this ELISA. The test resulted with an average level of 375 µg/mL (1.875 µg/mL measured directly from the calibrator curve), ranges from 275.6 – 505.2 µg/mL, Std Dev of 60.4 µg/mL).

We strongly recommend for each clinical laboratory to establish its own normal range by measuring EDTA plasma and/or serum with this kit.

Reference Range:

200 – 550 µg/mL (L.Thomas, 1982)

Additional Reference Ranges:

Pregnant woman

Samples of pregnant women were measured to have a 30-80% higher reference range than the control groups.

Liver diseases

According to Haughton et al., the reference range of patients with liver diseases is 35% lower than the one of healthy controls.

XIV. LIMITATION OF THE PROCEDURE

1. This assay requires serum or plasma sample for testing.
2. Serum or plasma samples from different species may show different matrix background.
3. For sample values greater than 12.5 µg/mL, it is recommended to re-assay samples with dilution (i.e. 1:10 or 1:100). The best dilution matrix is vitamin D free human serum.
4. Cell culture or tissue culture samples should be validated with total binding and other performance specifications before being used.
5. Severely hemolyzed samples, icteric or lipaemic sample should not be used

XV. QUALITY CONTROL

To assure the validity of the results, each assay should include adequate controls with known positive levels of VDBP. We recommend that all assays include the laboratory's own control samples in addition to those provided with this kit.

XVI. PERFORMANCE CHARACTERISTICS

Sensitivity

The analytical sensitivity (LLOD) of this VDBP ELISA as determined by the 2 times standard deviation below the mean of B₀ (B₀ - 2SD) on 16 duplicate determinations of calibrator 1 (B₀) and calibrator 2 with a concentration of 0.205µg/mL . The LLOD of this test is approximately 0.0637 µg/mL.

Precision

The intra-assay precision was validated by measuring three diluted 1:200 samples with 16 replicate determinations.

Sample 1:200	VDBP Value (µg/mL)	CV (%)
1	264.2	8.8
2	1008.6	8.1
3	1945.4	9.1

The inter-assay precision was validated by measuring two control levels in duplicate in 13 individual assays.

Sample	Mean VDBP Value (µg/mL)	CV (%)
1	184.8	8.7
2	1164.9	5.9

Linearity

Three samples were collected, diluted 1:200 and spiked with various amounts of VDBP, diluted with standard zero matrix and tested. The results of VDBP percent recovery value in VDBP concentration µg/mL are as follows:

DILUTION 1:200	VDBP VALUE (µg/mL)	RECOVERY %
Sample A	2030	-
1:2	1006	100
1:4	495.4	99
1:8	226.4	90
Sample B	750.4	-
1:2	402.8	107
1:4	208.6	111
1:8	90.6	97
Sample C	220.8	-
1:2	103.6	94
1:4	55.4	100
1:8	21.8	79

Spike Recovery

Three diluted 1:200 samples and three assay calibrators (0.8, 2.0 and 5 µg/mL) were combined at equal volumes and tested. The results are as follows:

DILUTION 1:200	VDBP VALUE (ng/mL)	RECOVERY %
Sample A	247.8	-
0.8 µg/mL	203.8	100
2.0 µg/mL	361	111
5.0 µg/mL	634.2	102
Sample B	297.2	-
0.8 µg/mL	234.6	103
2.0 µg/mL	403.2	115
5.0 µg/mL	662.2	102
Sample C	222	-
0.8 µg/mL	189.8	99
2.0 µg/mL	337.6	109
5.0 µg/mL	587.4	96

Interference

One positive sample is added with 5% volume of interference materials to reach a final concentration shown in the table below. All samples are tested in an assay in duplicate.

	Mean OD 450 nm		
	Additive	Amt Added (mg/mL)	Sample
1	Test Control	-	405.5
2	Bilirubin - L	0.4	433.0
3	Bilirubin - H	10	394.7
4	Test Control	-	476.2
5	Hb - L	0.26	461.0
6	Hb - H	6.5	478.2
7	Lipid - L	8	368.4
8	Lipid -H	200	456.3


XVII. WARRANTY

This product is warranted to perform as described in its labeling and literature when used in accordance with all instructions. Epitope Diagnostics, Inc. DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, and in no event shall Epitope Diagnostics, Inc. be liable for consequential damages. Replacement of the product or refund of the purchase price is the exclusive remedy for the purchaser. This warranty gives you specific legal rights and you may have other rights, which vary from state to state.












XVIII. REFERENCES

- Haddad JG¹. Plasma vitamin D-binding protein (Gc-globulin): multiple tasks. J Steroid Biochem Mol Biol. 1995 Jun;53(1-6):579-82.
- Carpenter TO¹, Zhang JH, Parra E, Ellis BK, Simpson C, Lee WM, Balko J, Fu L, Wong BY, Cole DE. Vitamin D binding protein is a key determinant of 25-hydroxyvitamin D levels in infants and toddlers. J Bone Miner Res. 2013 Jan;28(1):213-21. doi: 10.1002/jbmr.1735.

TECHNICAL ASSISTANCE AND CUSTOMER SERVICE
 For technical assistance or place an order, please contact Epitope Diagnostics, Inc. at (858) 693-7877 or fax to (858) 693-7678. www.epitopediagnostics.com

 **Epitope Diagnostics, Inc.**
 San Diego, CA 92121, USA

  MDSS GmbH
 Schiffgraben 41
 30175 Hannover,

 Manufacturer	 No. of tests
 Catalog Number	 Keep away from heat and direct sun light
 Concentrate	 Store at
 In Vitro Diagnostic Device	 Use by
 Read instructions before use	 Lot No.
 Authorized Representative In Europe	

VDBP ELISA: Condensed Assay Protocol

- 15 µl calibrators, controls and unknown 1:200 samples + 100 µL Tracer Antibody
 ↓ Incubate @RT for 60 min, static Wash 5x
- 100 µl HRP Substrate
 ↓ Incubate @RT for 20 min static
- 100 µl Stop Solution
 ↓ Immediately
- Read absorbance at 450 nm (4-parameter)
 within 10 minutes